



America

CERTIFICATE


No. U8 019612 0076 Rev. 00

Holder of Certificate: **Amann Girrbach AG**



Herrschaftswiesen 1
6842 Koblach
AUSTRIA

Certification Mark:



C US

Product: **Laboratory Equipment**

Model(s): **Ceramill Motion 2, Ceramill Motion 2 (4X),
Ceramill Motion 3**

Parameters:

Rated voltage:	100-230V
Rated frequency:	50-60Hz
Rated power consumption:	250W
Protection class:	I

Tested according to: UL 61010-1:2012/R:2019-07
CSA C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713256508

Date, 2022-06-09

Siemon

(Thorsten Siemon)